



DIAMOND DENTAL BURS USER MANUAL

CE
0086

Intended use

Produced for all kinds of carving and cutting operations which are used by dentists during medical treatments. Area of application These general instructions for use and safety recommendations apply to all products and have to be generally observed. Nonobservation of these instructions for use and safety recommendations increases the risk of injury and may impair the proper function of the instruments. Separate instructions of use are available on our web site www.bosphorus.com.tr for products that require more detailed information.

These take precedence over the general instructions.

Storage

Instruments in their original packaging, away from light and heat in a clean and dry environment. Do not store in the same room as solvents or chemicals.

1. Proper use

- The instruments have to be prepared prior to first use.
 - Make sure that only technically and hygienically perfect and cleaned power systems (hand pieces and contra-angles) are used.
 - Depending on their shank type, insert the instrument into the chuck as deeply as possible. Make sure that they are properly locked.
 - The instrument must be rotating at the desired speed before contact is made with the tissue or material.
 - Avoid jamming and using the instrument as a lever. Excessive contact pressure has to be avoided.
 - For eye protection wear safety glasses. Use appropriate respiratory protection (mouth and nose). In the dental laboratory, use appropriate suction unit.
 - Avoid contact with the instruments' working parts as this may increase the risk of injury.
- Make sure to provide sufficient cooling by means of air/water spray. Additional external cooling is required when using instruments with extra-long shanks or oversized working parts. Damaged or corroded instruments have to be discarded.

2. Recommended speeds

HEAD DIAMETER(ISO Ø (1/10 mm)	FG(Max. RPM)	HP(Max. RPM)
007-014	300,000	160,000
016-023	280,000	120,000
025-045	180,000	100,000
047-065	80,000	80,000
066-093	60,000	50,000

For optimum results when using Bosphorus branded burs, 50% of the maximum speed should be used..

3. Contact pressure

- Excessive contact pressure (> 2N) has to be avoided.
- In cutting instruments, this can lead to damage to the working part and to chips of the diamond as well as an excessive generation of heat.



4. Guideline on the number of times rotary instruments can be used

The below values are guidelines. The service life of instruments differ from these values as this depends on the application and /or the material treated

Diamond dental bur : 15x

The Reuse of disposable products is not permitted

5. Disposal

To prevent contamination, discard instruments in tight, puncture resistant containers.

6. Disinfection, cleaning and sterilisation

Those instruments that are supplied non-sterile have to be prepared prior to first use.

6.1. Manual reprocessing

The instruments are to be disinfected with appropriate cleaning and disinfecting agents for these product for recommendations for use immersion time concentration, rinsing,drying of cleaning abd disinfecting agents see instructions of the manufacturers of these agents. MAKE sure that the instruments do not come in contact with each other during ulasonic cleaning

6.2. Mechanical reprocessing

The instruments have to be treated with suitable detergents and disinfectants that are recommended for this purpose. Observe the instructions of use provided by the manufacturer.

Make sure that the instruments do not come in contact with each other during the cleaning in the ultrasonic bath. Inspect the clean instruments visually. Separate and discard damaged or blunt instruments. Thoroughly cleaned instruments are an essential condition for a successful sterilization

6.3. Sterilization

Make sure that the instruments are sterilized according to a validated procedure suitable for the sterilization of medical products. Successful sterilization also depends on the type of product to be sterilized, the packaging and the loading set-up of the sterilization device. The operator of medical products is responsible for seeing that proper treatment is carried out by qualified personnel, using the appropriate materials and suited equipment, as recommended by the Commission for Hospital Hygiene and Infectious.

Caution:Operative burs must be sterilized before each use to avoid potential transfer of infectious organisms.Autoclaving:

Dip the instruments in a corrosion inhibitor such as 1% sodium nitrite, prior to the sterilization cycle. In the unwrapped condition, the instrument can be sterilized by a pre-vacuum flash or gravity flash steam sterilization,utilizing a steam sterilizer. For the pre-vacuum Flash method, operate at full cycle with a dwell at 134°C minimum for 5 minutes.

7. Specific instructions for individual instrument types

- Avoid any contact with H2O2 (hydrogen peroxide). Make sure that the specified immersion times in the cleaning and disinfecting agents are not exceeded.

8. Potential effects of using worn instruments

- Discard any instruments with damaged or worn working parts as the use of damaged or worn instruments would have a negative effect on the work result.

- Damaged and deformed burs will cause the instrument

to vibrate and lead to poor preparation margins and rough surfaces.

- Please avoid excessive contact pressure during use as this may result in damage to the working part (Nicks on the blades, premature blunting and excessive heat generation).

- To avoid undesirable heat generation during preparation, make sure to provide sufficient cooling by means of air/water spray (at least 50 ml/min).

- Additional external cooling is required when using instruments with a total length of more than 22 mm or a head diameter exceeding 2,5 mm.

- Not observing the maximum permissible speed will result in an increased risk of injury.

- There is an increased risk of infection in reusable instruments which have not been properly reprocessed.

9. Safety and possible side effects

The above mentioned recommendations with respect to cooling, contact pressure, disinfection, cleaning and sterilization are to be strictly observed. The instruments should only be used for the intended application, as per the symbolic identification. Non-observance of these safety recommendations may lead to damage of the power system and/or injury, such as thermal necrosis, undesired preparation of tissue, damage to tissue or nerves, violation of the biological width, or infections. During preparation, some instruments may generate metallic abrasion which may lead to the presence of artifacts in MRI diagnostics.

10. Liability

It is the responsibility of the user to check the products prior to use to ensure that they are suited for the intended purpose. In case of contributory negligence by the user, Bosphorus partially or totally declines liability for all resulting damages, particularly if these are due to non-observance of our recommendations for use or warnings as well as inadvertent misuse by the user.

Store products out of children's reach.

For dental use only

Symbols of Label:

	MANUFACTURER
	CATALOGUE NUMBER
	BATCH CODE
	CAUTION
	Non-Sterile
	CONSULT INSTRUCTIONS FOR USE
	According to EN ISO 6360-1,2,3,5,6,7
	Defines type of operative burs

TEK METAL SAN. VE TIC. A.S.
Ikitelli Org.San.Bolgesi Demirciler
San.Sit C-2 Blok No:209 Ikitelli-
İstanbul - TURKEY

Tel: +902126715104 Fax: +902125497032
Web: www.bosphorus.com.tr
E-mail: info@bosphorus.com.tr

CE 0086



ELMAS DENTAL FREZ KULLANIM KILAVUZU

CE
0086

Kullanım Amacı

Tıbbi tedaviler sırasında dişçiler tarafından kullanılan her türlü oyma ve kesme operasyonları için üretilmiştir. Uygulama Alanı Kullanım için genel talimatlar ve güvenlik tavisileri tüm ürünleri kapsamaktadır ve genel olarak izlenmelidir. Kullanım talimatları ve güvenlik tavisilerini izlenmemesi yaralanma riskini artırılmaktır ve aletlerin düzgün kullanılmamasına neden olabilmektedir. Detaylı bilgi gerektiren ürünler için ayrı kullanım talimatları web sitemizde www.bosphorus.com.tr mevcuttur. Bunlar genel talimatların üzerindedir.

Depolama

Frezleri orijinal paketlerinde, ışık veya ısidan uzakta temiz ve kurutmadan tutunuz. Solventler veya kimyasallara aynı odada tutmayın.

1. Amaca Uygun Kullanım

- Frezler uygun ilk kullanım için hazırlanmalıdır.
- Teknik ve hijyenik olarak mükemmel ve temiz güç sistemlerinin kullanılıldığından emin olunuz.
- Saft tipine göre, aleti kilitleme tertibatının içinde olabildiğince derine koynuz. Düzgüne kilitlendiklerinden emin olunuz.
- Alet doku veya materyal ile temas etmeden önce istenilen hızda dönelmeli olmalıdır.
- Sıkışmadan ve aleti tesviye olarak kullanmaktan kaçınınız. Asırı temas basınıcını kazınmalıdır.
- Gör koruması için güvenlik gözlüğü takınız. Uygun aspiratör koruması (ağız ve burun) kullanınız. Diş laboratuvarında uygun vakum birimi kullanınız.
- Yaralanma riskini arıtarabileceğinden aletin çalışan kısımları ile temastan kaçınınız.
- Hava/su spreyi ile yeterli soğutmanın sağlığından emin olunuz. Ek diş soğutma, aletle ekstra uzun sürt veya aşırı büyük çalışan kısımlar kullanıldığından gereklidir. Hasarlı veya aşınmış aletler atılmamalıdır.

2. Tavsiye edilen hızlar

KAFADİP(ISO Ø (1/10 mm)	FG(Max. RPM)	HP(Max. RPM)
007-014	300,000	160,000
016-023	280,000	120,000
025-045	180,000	100,000
047-065	80,000	80,000
066-093	60,000	50,000

Bosphorus markalı frezleri kullanırken optimum sonuçlar için maksimum hızın% 50'si kullanılmalıdır.

3. Temas Basıncı

Aşırı temas basıncından (>2N) kaçınılmaktadır.

- Frezlerde, çalışan kısımlarda hasara ve aşırı ısınma yanında elmasların neden olabilir.

4. Dönen aletlerin kullanım sayıları

Aşağıdaki değerler kilavuz niteliğindedir. Cihazların hizmet ömrü, uygulamaya ve/veya işlenen malzemeye bağlı olarak bu değerlerden farklıdır
Elmas dış frez : 15x
Tek kullanımlık ürünlerin yeniden kullanımına izin verilmez

- İzin verilen maksimum hız uygulaması artan yaralanma riskini doğuracaktır.
- Düzenin şekilde yeniden işlenmeyen tekrar kullanılabilen aletlerde artan enfeksiyon riski vardır.

9. Güvenlik ve muhtemel yan etkiler

Yukarı bahsedilen sóğutma, temas basıncı, dezenfeksiyon, temizlik ve sterilizasyon tavisiyle kesinlikle uygulanmalıdır. Aletler, yalnızca sembolik belirtimlere göre planlanan uygulama için kullanılmıştır. Bu güvenlik tavisilerinin uygulanması güç sisteminde hasar ve/veya termal nekroz, dokunun istenmeyen hazırlığı, doku veya sinir hasarı, bıyo lojik gelişimi ihali veya enfeksiyonlar gibi yaralanmaya neden olabilir. Hazırlama sırasında bazı aletler metalik aşınma oluşturabilir ve bu da MRI tanda kalıntı varlığına neden olabilir.

10. Yükümlülük

Kullanımdan önce planlanan amaç için uygunluğun kontrolü kullanıcının sorumluluğudur. Kullanıcı tarafından müşterek kusur durumunda, Bosphorus, özellikle kullanım tavisilerinin veya uyarıların izlenmemesi, kullanıcı tarafından istemeden yanlış kullanım sonucunda ise kısmen veya tamamen olabileceği hasarlardaki yükümlülüğü red eder. Çocukların ulaşamayacağı yerlerde saklayınız. Yalnızca dışçılık kullanımı içindir.

Etket Sembollerı:

	ÜRETİCİ
	KATALOG NUMARASI
	LOT NO
	DİKKAT
	STERİL DEĞİL
	KULLANMA KILA VUZUNU İNCELEYİN
	EN ISO 6360-1,2,3,5,6,7 YA GÖRE
	FREZLERİN TİPİNİ BELİRTİR

TEK METAL SAN. VE TIC. A.S.
İkitelli Org.San.Bolgesi Demirciler
San.Sit C-2 Blok No:209 İkitelli-
İstanbul - TURKİYE
Tel: +902126715104 Fax: +902125497032
Web: www.bosphorus.com.tr
E-mail: info@bosphorus.com.tr

CE 0086